



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Chant Engineering Co. Inc.

59 Industrial Drive, New Britain, PA 18901

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017 & Meets the Requirements of ANSI/NCSL Z540.1-1994

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Force Calibration *(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

January 16, 2016

Issue Date:

January 29, 2022

Expiration Date:

April 30, 2024

Accreditation No.:

87679

Certificate No.:

L22-24

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Chant Engineering Co. Inc.

59 Industrial Drive, New Britain, PA 18901
Contact Name: Jules Dirienzo Phone: 215-230-4260

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure: Force and Tension Source and Measure ^{FO}	100 lbf to 5 000 lbf (444 N to 22 241 N)	0.09 % of reading	8605-01 Tension Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500
	1 000 lbf to 60 000 lbf (4 448 N to 266 893 N)	0.48 % of reading	8603-02 Tension Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500
	11 332 lbf to 350 000 lbf (50 407 N to 1 556 877 N)	0.11 % of reading	8605-03 Tension Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500
	98 000 lbf to 1 500 000 lbf (435 925 N to 6 672 332 N)	0.11 % of reading	8604-01T Tension Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500
	202 000 lbf to 1 650 000 lbf (898 540 N to 7 339 565 N)	0.28 % of reading	8604-01C Compression Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500
	1 000 lbf to 60 000 lbf (4 448 N to 226 893 N)	0.75 % of reading	8607-01T Tension Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500



Certificate of Accreditation: Supplement

Chant Engineering Co. Inc.

59 Industrial Drive, New Britain, PA 18901
Contact Name: Jules Dirienzo Phone: 215-230-4260

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Equipment to Measure: Force and Tension Source and Measure ^{FO}	1 000 lbf to 65 000 lbf (4 448 N to 289 134 N)	0.09 % of reading	8607-01C Compression Load cell with indicator WI: QSP-QC-02 IAW ASTM-E4, AS2193, ISO-7500

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.